

# Avery Weigh-Tronix

## Calibration Certificate

Avery Weigh-Tronix  
 Foundry Lane, Smethwick  
 West Midlands, B66 2LP  
[www.averyweigh-tronix.com](http://www.averyweigh-tronix.com)  
**Tel:** 0845 900 22 44



### Customer Details

Name : Highlander International Ltd  
 Address : Linwood Avenue  
 Collage Milton  
 East Kilbride  
 G74 5NR  
 Contact Name : Brian Bingham  
 Telephone No. 01355524215

### Linearity / Accuracy Test

Test Point	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
0 kg			0 kg	0 kg
4,000 kg			4,000 kg	0 kg
8,000 kg			8,000 kg	0 kg
12,000 kg			12,000 kg	0 kg
16,000 kg			16,000 kg	0 kg
22,700 kg			22,700 kg	0 kg
31,580 kg			31,580 kg	0 kg
39,580 kg			39,580 kg	0 kg
46,280 kg			46,280 kg	0 kg

**Certificate No. 920982047/0101**

**Date of Calibration :** 24/10/2016  
**Next Calibration Due :** Oct 2017

### Agreement No.

Location : EAST KILBRIDE  
 Department :  
 Plant Number :  
 Manufacturer : Avery Weigh-Tronix  
 Model : ZM305  
 Serial No : 163050050  
 Capacity : 50000 kg  
 Resolution : 20kg  
 Tolerance : +/- 60kg  
 Calibration Frequency :

### Off Centre 2 3 or End -Middle-End Loading Test 1 4 1-2-3

Test Point	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
10,000 kg			10,000 kg	0 kg
10,000 kg			10,000 kg	0 kg
10,000 kg			10,000 kg	0 kg
10,000 kg			10,000 kg	0 kg

### Test Weight Details

Weight Set No. & Certificate No.

W004: A0264

Test weights and equipment are traceable to national standards.

### Repeatability Test

Test Point	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
39,580 kg			39,580 kg	0 kg
39,580 kg			39,580 kg	0 kg
39,580 kg			39,580 kg	0 kg

This certificate complies with the United Kingdom Weighing Federation Calibration Code of Practice for non-automatic weighing instruments.

### Comments

Calibration to Weighbridge. Make-up weight used  
 No as found readings: New installation

### Tolerances

Unless otherwise agreed tolerances will be those permitted by current Weights and Measures Regulations for Class III machines as follows:

0-500 divisions +/- 1 divn.  
 >500-2000 divisions +/- 2 divn.  
 >2000 divisions +/- 3 divn.

For key calibration information please login to <https://averyweigh-tronix.com>

Technicians Name: Williamson W

Customer Name: Margaret Bingham

Technicians Signature:

Customer Signature: